

## Material Safety Data Sheet – InSight HCT Meter

### 1. IDENTIFICATION OF PRODUCT AND SUPPLIER

**Product Name:** InSight HCT Meter  
**Chemical Family or Product Type:** *In-vitro* diagnostic  
**Supplier:** Woodley Equipment Company Ltd.  
Old Station Park Buildings, St. John Street,  
Horwich, Bolton, Lancashire, BL6 7NY, UK  
**Tel:** +44 (0) 1204 669033  
**Email:** [sales@woodleyequipment.com](mailto:sales@woodleyequipment.com)

### 2. HEALTH HAZARD INFORMATION

**Hazard Designation:** Void  
**Information Pertaining to Particular Dangers for Man and Environment:** Void

**Classification System:**

The classification is in line with current EC lists. It is expanded, however, by information from technical literature and by information furnished by supplier companies.

### 3. SPECIAL SAFETY MEASURES

All conduction emission and radiation emission passed EN 61326.

### 4. STORAGE ENVIRONMENT

**Environment Temperature:** -20°C~55 °C

**Comparative Humidity:** ≤95% RH

The packed instrument should be stored in a ventilated room. Do not store the instrument with toxic or corrosive materials.

### 5. FIRE-FIGHTING DATA

**Suitable Extinguishing Agents:** Use extinguishing media suitable for surrounding fire

**Protective Equipment:** No special measures required.

### 6. TRANSPORT INFORMATION

**Land Transport ADR/RID and GGVS/GGVE (cross-border/domestic) ADR/RID-GGVS/E Class:** Not regulated

**Conditions to Avoid:** Excessive heat or cold, direct light, high humidity.

InSight HCT Meter is a precise instrument. Transport it with utmost care and avoid moisture, sunlight and collision.

**Transportation Condition:** -40°C~55 °C; RH: ≤95% RH

### 7. USING TIME LIMIT

The life span of the instrument design is 8 years.

#### **8. EXPOSURE CONTROL**

Fresh whole blood

**Handling:** Keep samples away from direct sunlight

**Protective Clothing:** Wear protective gloves to avoid exposure to blood samples

**Reagent Strips:**

Exposing the reagent strips to air for a long time will deteriorate the compositions of the reagent strips.

Do not touch the reagent pads on the reagent strips.

#### **9. USING ENVIRONMENT**

**Operating Temperature:** 5 °C~ 40 °C;

**Measuring Temperature:** 15 °C~ 30 °C;

**Comparative Humidity:** ≤80% RH; Barometric:  
80kpa ~ 106kpa

#### **10. SAFETY POWER**

**Voltage:** DC6V (x2 CR2032 Battery)

#### **11. DISPOSAL CONSIDERATIONS**

Wear protective gloves to avoid exposure to pathogenic microbes.

Dispose of used samples, reagent strips and gloves according to local regulations.

#### **12. HANDLING**

**The instrument must be handled with care. Careless handling may disturb internal calibration and/or cause other malfunctions. An error code will appear on the display if there is something wrong with the instrument and the error code will also be printed out if the printer is attached. The table below presents the various error codes, along with possible causes and recommended remedies to quickly and easily solve the problems. If the error code persists, contact Woodley Equipment Company.**

**WHILE WOODLEY EQUIPMENT COMPANY LTD. BELIEVES THE INFORMATION CONTAINED HEREIN IS VALID AND ACCURATE, WOODLEY EQUIPMENT MAKE NO WARRANTY OR REPRESENTATION AS TO ITS VALIDITY OR ACCURACY. WOODLEY EQUIPMENT SHOULD NOT BE LIABLE OR OTHERWISE RESPONSIBLE IN ANY WAY FOR USE OF EITHER THIS INFORMATION OR MATERIA TO WHICH IT APPLIES.**